
































Symbols Glossary









SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	ISO 15223-1, Clause 5.1.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Catalogue number	Indicates the manufacturer's catalogue number so that the medical device can be identified.
	ISO 15223-1, Clause 5.1.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Manufacturer	Indicates the medical device manufacturer.
	ISO 15223-1, Clause 5.2.3	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Sterilized using ethylene oxide	Indicates a medical device that has been sterilized using ethylene oxide.
	ISO 15223-1, Clause 5.2.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Do not resterilize	Indicates a medical device that is not to be resterilized.
	ISO 15223-1, Clause 5.2.7	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Non sterile	Indicates a medical device that has not been subjected to a sterilization process.
	ISO 15223-1, Clause 5.2.8	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied .	Do not use if package is damaged	Indicates a medical device that should not be used if the package has been damaged or opened.









SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	ISO 15223-1, Clause 5.4.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Do not reuse	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.
	ISO 15223-1, Clause 5.6.3	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Non-pyrogenic	Indicates a medical device that is non-pyrogenic.
	ISO 15223-1, Clause 5.4.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Caution	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.
	ISO 15223-1, Clause 5.1.5	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified.
	ISO 15223-1, Clause 5.1.7	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Serial number	Indicates the manufacturer's serial number so that a specific medical device can be identified.
	ISO 15223-1, Clause 5.1.3	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Date of manufacture	Indicates the date when the medical device was manufactured.
	ISO 15223-1, Clause 5.1.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Use-by date	Indicates the date after which the medical device is not to be used.





SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	SO 15223-1, Clause 5.1.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Authorized European representative	Indicates the Authorized representative in the European Community.
	ISO 15223-1, Clause 5.5.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	In vitro diagnostic medical device	Indicates a medical device that is intended to be used as an in vitro diagnostic medical device.
	ISO 15223-1, Clause 5.4.3	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied	Consult instructions for use	Indicates the need for the user to consult the instructions for use.
	IEC 60601-1, Table D.2, Symbol 10	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.	Follow instructions for use	Refer to instruction manual/booklet.
	ISO 7010, M002	Graphical symbols -- Safety Colours and Safety signs - Registered safety signs		
	IEC 60601-1, Table D.2, Symbol 2	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.	General Warning Sign	To signify a general warning
	ISO 7010, W001	Graphical symbols -- Safety Colours and Safety signs - Registered safety signs		
	EN 15986, Clause 4.2	Symbol for use in the labelling of medical devices. Requirements for labelling of medical devices containing phthalates.	DEHP di(2-ethylhexyl) phthalate	Indicates presence of di(2-ethylhexyl) Phthalate (DEHP).

SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	21 CFR 801.15(c)(1)(i)F	Labeling-Medical devices; prominence of required label statements.	Prescription only	Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.
	21 CFR 801.109	Labeling-Prescription devices.		
	EN 50419	Marking of Electrical and Electronic Equipment in accordance with Article 11(2) of Directive 2002/96/EC (WEEE).	Recycle: Electronic Equipment	Do not throw in trash.
	ISO 15223-1, Clause 5.3.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Fragile, handle with care	Indicates a medical device that can be broken or damaged if not handled carefully.
	ISO 15223-1, Clause 5.3.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Keep dry	Indicates a medical device that needs to be protected from moisture.
	ISO 15223-1, Clause 5.3.5	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Lower limit of temperature	Indicates the lower limit of temperature to which the medical device can be safely exposed.
	ISO 15223-1, Clause 5.3.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Upper limit of temperature	Indicates the upper limit of temperature to which the medical device can be safely exposed.
	ISO 15223-1, Clause 5.3.7	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Temperature limit	Indicates the temperature limits to which the medical device can be safely exposed.

SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	ISO 15223-1, Clause 5.3.8	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Humidity limitation	Indicates the range of humidity to which the medical device can be safely exposed.
	ISO 15223-1, Clause 5.3.9	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Atmospheric Pressure	Indicates the range of atmospheric pressure to which the medical device can be safely exposed.
IPX1 IPX5	IEC 60529	Degrees of Protection Provided by Enclosures (IP Code).	Degree of Ingress Protection Provided by Enclosure	Degree of protection against ingress of water and particulate matter
	IEC 60601-1, Table D.1, Symbol 8	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.	Equipotentiality	To identify the terminals which, when connected together, bring the various parts of an equipment or of a system to the same potential, not necessarily being the earth (ground) potential e.g. for local bonding.
	IEC 60417-5021	Graphical Symbols for Use on Equipment		
	IEC 60417-5335	Graphical Symbols for Use on Equipment	Type CF applied part	To identify a type CF applied part complying with IEC 60601-1
	IEC 60417-5007, 5008	Graphical Symbols for Use on Equipment	"ON" (power) "OFF" (power)	To indicate connection to or from the mains, at least for mains switches or their positions, and all those cases where safety is involved.

SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	IEC 60417-5010	Graphical Symbols for Use on Equipment	"ON"/"OFF" (push-push)	To indicate connection to or disconnection from the mains, at least for mains switches or their positions, and all those cases where safety is involved.
	IEC 60417-5019	Graphical Symbols for Use on Equipment	Protective earth; protective ground	To identify any terminal which is intended for connection to an external conductor for protection against electric shock in case of a fault, or the terminal of a protective earth (ground) electrode.
	IEC 60417-5034	Graphical Symbols for Use on Equipment	Input	To indicate an input terminal.
	IEC 60417-5035	Graphical Symbols for Use on Equipment	Output	To indicate an output terminal.
	IEC 60417-5036	Graphical Symbols for Use on Equipment	Dangerous voltage	To indicate hazards arising from dangerous voltages.
	IEC 60417-5236	Graphical Symbols for Use on Equipment	Draining	To identify the draining control.
	IEC 60417-5448	Graphical Symbols for Use on Equipment	Input/output	To identify a combined input/output connector or mode.
	IEC 60417-5795	Graphical Symbols for Use on Equipment	ECG triggered display	To identify the control or the indicator for the display of electrocardiogram (ECG) triggered sequences.

SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	IEC 60417-5850	Graphical Symbols for Use on Equipment	Serial interface	To identify a connector for a serial data connection.
	IEC 60417-6024	Graphical Symbols for Use on Equipment	Inlet for liquid	To identify inlet connections for the cooling liquid.
	IEC 60417-6025	Graphical Symbols for Use on Equipment	Outlet for liquid	To identify outlet connections for the cooling liquid.
	IEC 60417-6046	Graphical Symbols for Use on Equipment	Rated power input, d.c.	To indicate a d.c. rated power input.
	ISO 7000 - 0623	Graphical Symbols for Use on Equipment	This way up	To indicate correct upright position of the transport package.
	ISO 7000 - 3500	Graphical Symbols for Use on Equipment	Electronic instructions for use	To indicate on product or product packaging that relevant information for use of the product is available in electronic form rather than, or in addition to, printed paper form.
	Regulation (EU) 2017/745, Article 2 (43), Annex V	Medical Devices Regulation	CE Marking	Indicates that a device is in conformity with the applicable requirements set out in the Medical Device Regulation and other applicable Union harmonisation legislation providing for its affixing.
	N/A	N/A	USB Port	Identify USB (Universal Serial Bus) connection port.

SYMBOL	STANDARD REFERENCE	STANDARD TITLE	SYMBOL TITLE	EXPLANATORY TEXT
	N/A	N/A	Lithium Battery Class 9	Lithium ion batteries contained in equipment or Lithium ion batteries packed with equipment.
	N/A	N/A	N/A	Lithium Battery Caution. Do not load or transport package if damaged.
	N/A	N/A	N/A	Cargo Aircraft Only, Forbidden in Passenger Aircraft.
	N/A	N/A	N/A	Indicates quantity.